



Ref. Certif. No.

DE 2-036051

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Portable Tourniquet System
Name and address of the applicant	Delfi Medical Innovations, Inc 1099 8th Ave, West, Suite 106 Vancouver BC V6H 1C3, Canada
Name and address of the manufacturer	Delfi Medical Innovations, Inc 1099 8th Ave, West, Suite 106 Vancouver BC V6H 1C3, Canada
Name and address of the factory	Delfi Medical Innovations, Inc 1099 8th Ave, West, Suite 106 Vancouver BC V6H 1C3, Canada
Ratings and principal characteristics	external power supply: AC 100-240V, 50/60Hz; 2.3A; Class I; Output: DC 12V, 8.3A
Trademark (if any)	Delfi
Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	PTSii (9-2200-001, 60-1000-101-00), PTS for BFR (9-2200-001BFR)
Additional information (if necessary may also be reported on page 2)	Replaces CB Certificate US-TUVR-12175 due to change of NCB.
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005+A1 IEC 60601-1-6:2010+A1 IEC 60601-1-8:2006+A1 for national differences see test report
As shown in the Test Report Ref. No. which forms part of this Certificate	31690047 009

This CB Test Certificate is issued by the National Certification Body



TÜV Rheinland LGA Products GmbH
Tillystr. 2, 90431 Nürnberg, Germany
Phone + 49 221 806-1371
Fax + 49 221 806-3935
Mail: cert-validity@de.tuv.com
Web : www.tuv.com

Date: 2022-06-28

Signature:

Dipl.-Ing. M. Glagla